



TWD Scientific, LLC
10555 86th Avenue
Pleasant Prairie, WI 53158

888.323.3585 TEL | 262.605.3262 FAX

www.twdtradewinds.com

PART #:	DESCRIPTION:	LOT #:
DF8988P-S	DISPOSABLE REINFORCED NYLON STERILE WHITE FORCEPS WITH BLUNT TIP. FORCEPS ARE INDIVIDUALLY PEEL-PACKED.	201504240201
EXPIRATION DATE (MONTH/DAY/YEAR): 04/23/2020		

CERTIFICATE OF STERILIZATION

The validation and routine control of the sterilization process are carried out in accordance with the following standards:

ISO 11135: Sterilization of health care products – Ethylene oxide – Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices

Provided appropriately stored, products remain sterile for at least 5 years from the sterilization date.

CERTIFICATE OF CONFORMITY

We hereby certify that product mention above meets the specifications stated.

Product: Sterile Forceps

Lot #: 201504240201

Expiration: 04/23/2020