



**TWD TradeWinds, Inc.**  
**LakeView Corporate Park**  
**10555 86<sup>th</sup> Ave. Pleasant Prairie, WI 53158 USA**

|                     |  |              |                              |
|---------------------|--|--------------|------------------------------|
| <b>Date</b>         | <b>08/01/11</b>  | <b>From</b>  | <b>TWD TradeWinds, Inc.</b>  |
| <b>PART #:</b>      | <b>DF8988P-S</b>   | <b>Phone</b> | <b>888.323.3585</b>          |
| <b>DESCRIPTION:</b> | <b>White disposable sterile forceps made of reinforced nylon with a blunt tip.</b> | <b>Fax:</b>  | <b>262.605.3262</b>          |
|                     |  | <b>Web:</b>  | <b>www.twdtradewinds.com</b> |

### **Certificate of Sterilization**

The validation and routine control of the sterilization procedure are carried out in accordance with the following standards:

ISO 11135: Sterilization of health care products -- Ethylene oxide --  
 Requirements for development, validation and routine control of a  
 sterilization process for medical devices

Provided appropriate storage products remain sterile for at least 5 years from sterilization date. The bio monitors for the sterilization in accordance with ISO11135 have shown no bacteria growth. The sterilization of enclosed listed products has been approved.

### **Certificate of Conformance**

We hereby certify that below mentioned products meet the specifications stated.

**Product:** Sterile Forceps  
**Lot #:** 201106100301  
**Expiration:** 06/09/2016