



TWD TradeWinds, Inc.
LakeView Corporate Park
10555 86th Ave. Pleasant Prairie, WI 53158 USA

Date	5/13/13	From	TWD TradeWinds, Inc.
PART #:	DF8988P-S	Phone	888.323.3585
DESCRIPTION:	White disposable sterile forceps made of reinforced nylon with a blunt tip.	Fax:	262.605.3262
		Web:	www.twdtradewinds.com

Certificate of Sterilization

The validation and routine control of the sterilization procedure are carried out in accordance with the following standards:

ISO 11135: Sterilization of health care products -- Ethylene oxide --
 Requirements for development, validation and routine control of a
 sterilization process for medical devices

Provided appropriate storage products remain sterile for at least 5 years from sterilization date. The bio monitors for the sterilization in accordance with ISO11135 have shown no bacteria growth. The sterilization of enclosed listed products has been approved.

Certificate of Conformance

We hereby certify that below mentioned products meet the specifications stated.

Product: Sterile Forceps
Lot #: 201303200201
Expiration: 03/19/2018