



TWD Scientific, LLC  
10555 86<sup>th</sup> Avenue  
Pleasant Prairie, WI 53158

888.323.3585 TEL | 262.605.3262 FAX

[www.twdtradewinds.com](http://www.twdtradewinds.com)

PART #:	DESCRIPTION:	LOT #:
DF8988P-S	DISPOSABLE REINFORCED NYLON STERILE WHITE FORCEPS WITH BLUNT TIP. FORCEPS ARE INDIVIDUALLY PEEL-PACKED.	201404040201
EXPIRATION DATE (DAY/MONTH/YEAR): 04/03/2019		

## CERTIFICATE OF STERILIZATION

The validation and routine control of the sterilization process are carried out in accordance with the following standards:

ISO 11135: Sterilization of health care products – Ethylene oxide – Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices

Provided appropriately stored, products remain sterile for at least 5 years from the sterilization date.

## CERTIFICATE OF CONFORMITY

We hereby certify that product mention above meets the specifications stated.

Product: Sterile Forceps

Lot #: 201404040201

Expiration: 04/03/2019